6640 185th Ave. NE, Redmond, WA 98052 T.425.895.8617, F.425.702.9358



CERTIFICATION CONCERNING DESIGN AND CONSTRUCTION OF ELECTRONIC SPEED MEASURING DEVICES

IRLJ RULE 6.6 EFFECTIVE 1/3/2006

I Clifford W. Nahm, do certify under penalty of perjury, under the laws of the state of Washington as follows:

I am employed with Cascade Engineering Services, Inc. (CES) Metrology and Electronics Repair Services, as the Metrology Lab Manager. I have been employed in such a capacity since 1996. Part of my duties includes supervising the maintenance and repair of all doppler radar and laser speed measuring devices (SMD's) used by CLE ELUM POLICE DEPARTMENT.

All SMD's currently used by CLE ELUM POLICE DEPARTMENT are listed in Exhibit "A".

I maintain the following qualifications with respect to SMD(s): United States Navy trained electronics technician specializing in Precision Measurement Electronics (PMEL) C School & Systems/Communications/Radar A School. I have commercial experience in electronics and in the repair and calibration of Doppler and Lidar SMD's dating back to 2006. I am experienced and competent in the principles and fundamental requirements of calibration from DC to Microwave frequencies.

The CES laboratory maintains manuals for all of the SMD's listed in Exhibit "A". I am personally familiar with those manuals and how each of the SMD's are designed and operated. On the date indicated in Exhibit "A" testing of the SMD's was performed using CES procedures under the direction of an authorized SMD expert. The results were evaluated and certified to meet or exceed existing performance standards and entered into the CES certification management database. CES laboratory maintains a testing and certification program that requires each SMD to be tested and certified for accuracy at least once every two years.

The CES laboratory tests all Doppler SMD's used by CLE ELUM POLICE DEPARTMENT, as recommended by the manufacturer, as follows: The Vocar HR, handheld Radar certification system is used to simulate speeds at 5 mph increments from 20 mph to 140 mph to verify accuracy in stationary and moving mode. Measurements are taken of the SMD transmit frequency, antenna/receiver sensitivity and any accompanying tuning forks are also tested for accuracy. All other operational functions of the SMD system are then tested for proper performance.

The Laser SMD's transmit a series of highly focused light wave pulses each time the trigger is pulled and utilizes two laws of physics; time and distance (i.e. 3.5 feet in diameter at 1000 feet). Since the speed of light is a known fixed value, the distance of the target is determined by calculating how long it takes for the signal to travel to the target and back. This series of measurements allows the SMD to calculate the speed of the target by measuring the distance the signal took to travel to the target and back. The displayed speed is accurate to within plus (+) or minus (-) one (1) mile per hour.

The CES laboratory tests all Laser / Lidar SMD(s) used by CLE ELUM POLICE DEPARTMENT, as recommended by the manufacturer, as follows: The Laser Speed Measurement Simulator (LSMS) is utilized to simulate a moving target. This is accomplished by detecting the optical output pulses of the laser device and generating artificial return pulses. Different speed values and ranges are simulated by varying the time delays between the input pulses and the return pulses. The LSMS consists of a Digital Delay Generator (DDG), and an optical interface unit. The DDG produces precise time delays. The optical interface unit converts the optical energy of the laser instrument into electrical signals which are supplied to the DDG. The optical interface unit also converts the electrical signals received from the DDG into optical energy which is then transmitted to the Lidar. The Lidar's output power is tested using an Ophir Nova Display, with a PD300-SH power head.

On the date indicated in Exhibit "A", each SMD was tested by a trained technician listed therein and under my direction. All Technicians listed on Exhibit "A" received training in the proper use and operation of SMD test equipment and performance testing procedures used to test Laser and Doppler SMDs. After successfully completing training the technician is certified by myself and receives authorization allowing them to enter the results from the tests into the certificate management database. Individual Performance and Certification tests are entered into the certificate management database under the penalty of perjury by entering an authorized user id and password to authenticate it.

Based upon my education, training, and experience and my knowledge of the SMD's listed above, it is my opinion that each of these electronic pieces of equipment is so designed and constructed as to accurately employ the Doppler effect in such a manner that it will give accurate measurements of the speed of motor vehicles when properly calibrated and operated by a trained operator or, in the case of the laser SMDs, each of these pieces of equipment is so designed and constructed as to accurately employ measurement techniques based on the velocity of light in such a manner that it will give accurate measurements of the speed of motor vehicles when properly calibrated and operated by a trained operator.

Exhibit "A" derives information from the certificate management database. See Exhibit "A" for details about individual SMD certifications.

State of Washington County of King

Signed or attested before me on

3-23-2021

by Clifford W. Nahm

I have satisfactory evidence that the person

described in this document:

(a) is personally known to me; OR (b) is identified upon oath oraffirmation

of credible witness personally know to me; OR

(c) is identified on the basis of

Indentification documents.

William Quoc Ang

Notary Public in and for the State of Washington,

Residing in Seattle, WA

My appointment expires January 29, 2022

Certified by: Clifford W. Nahm

Place: Redmond, WA



This agency, CLE ELUM POLICE DEPARTMENT currently utilizes the following Doppler SMD(s):

APPLIED CONCEPTS INC. manufacturer's the following SMD(s):

I.D./Serial Number	Model Number	Antenna 1 S/N	1 S/N Antenna 2 S/N T.F. 1 S/N T.F. 2 S/N Cal. Date Cal. Interval Due Date	T.F. 1 S/N	T.F. 2 S/N	Cal. Date	Cal. Interval	Due Date	Technician
DC106052	STALKER DUAL	KCO52707	N/A	FA286716	FB394845 (03/19/2021	12 MONTHS 0	FA286716 FB394845 03/19/2021 12 MONTHS 03/19/2022 JOHN R GRAY	IN R GRAY

KUSTOM SIGNALS, INC. manufacturer's the following SMD(s):

I.D./Serial Number	Model Number Antenna 1 S/N	Antenna 1 S/N	Antenna 2 S/N	T.F. 1 S/N	T.F. 2 S/N	Cal. Date	Antenna 2 S/N T.F. 1 S/N T.F. 2 S/N Cal. Date Cal. Interval Due Date	Technician
XE15018	EAGLE II	DE16734	DE27084	32829	34550	03/19/2021	03/19/2021 12 MONTHS 03/19/2022 JOHN R GRAY	HN R GRAY
XE15010	EAGLE II	DE26973	DE27001	61207	61442	03/19/2021	03/19/2021 12 MONTHS 03/19/2022 JOHN R GRAY	HN R GRAY
FF12409	FALCON	HANDHELD	N/A	26500	N/A	03/19/2021	03/19/2021 12 MONTHS 03/19/2022 JOHN R GRAY	HN R GRAY
XE08339	GOLDEN EAGLE II DE27085	DE27085	N/A	39205	41098	03/19/2021	03/19/2021 12 MONTHS 03/19/2022 JOHN R GRAY	HN R GRAY
XE08340	GOLDEN EAGLE II DE16711	DE16711	N/A	32821	34576	03/19/2021	03/19/2021 12 MONTHS 03/19/2022 JOHN R GRAY	HN R GRAY
XE14969	GOLDEN EAGLE II DE27103	DE27103	DE27064	39235	41070	03/19/2021	03/19/2021 12 MONTHS 03/19/2022 JOHN R GRAY	HN R GRAY
XE15012	GOLDEN EAGLE II DE16732	DE16732	N/A	39203	41093	03/19/2021	03/19/2021 12 MONTHS 03/19/2022 JOHN R GRAY	HN R GRAY
AA11249	HR-12	HANDHELD	N/A	292168	24719	03/19/2021	03/19/2021 12 MONTHS 03/19/2022 JOHN R GRAY	HN R GRAY
KK38300	TROOPER	CC39303	N/A	26717	25548	03/19/2021	03/19/2021 12 MONTHS 03/19/2022 JOHN R GRAY	HN R GRAY

MPH INDUSTRIES manufacturer's the following SMD(s):

in in	Condition of the second
Technicia	N R GRAY
Due Date	03/19/2021 12 MONTHS 03/19/2022 JOHN R GRAY
Cal. Interval	2 MONTHS 0
Cal. Date	03/19/2021 1
1 T.F. 2 S/N	854244
1 T.F. 1 S/N	854221
1 S/N Antenna 2 S/N T.F. 1 S/N T.F. 2 S/N Cal. Date Cal. Interval Due Date	N/A
ber Antenna 1 S/N	PYT831002137 N/A
Model Number	PYTHON III
I.D./Serial Number	PYT846001765

2 of 3







ONSITE DATE: Friday, March 19, 2021

CURRENT PO:

CLE ELUM POLICE DEPT

Asset / ID Number	Asset / ID Number MANUFACTURER	MODEL NUMBER	DESCRIPTION	DEPARTMENT	RESU
AA11249	KIISTOM SIGNALS INC	HR-12	SPEED MEASURING DEVICE	N/A	PAS
DC106052	APPLIED CONCEPTS INC	STALKER DUAL	SPEED MEASURING DEVICE	N/A	PAS:
FF12409	KUSTOM SIGNALS, INC.	FALCON	SPEED MEASURING DEVICE	N/A	PAS:
KK38300	KUSTOM SIGNALS, INC.	TROOPER	SPEED MEASURING DEVICE	N/A	PAS
PYT846001765	MPH INDUSTRIES	PYTHON III	SPEED MEASURING DEVICE	N/A	PAS
XEORASO	KUSTOM SIGNALS INC.	GOLDEN EAGLE II	SPEED MEASURING DEVICE	N/A	PAS
XE08340	KIISTOM SIGNAL S. INC.		SPEED MEASURING DEVICE	N/A	PAS:
XF14969	KUSTOM SIGNALS, INC.		SPEED MEASURING DEVICE	N/A	PAS:
XE15010	KIISTOM SIGNALS INC.	EAGLE II	SPEED MEASURING DEVICE	N/A	PAS
XF15012	KUSTOM SIGNALS, INC.	GOLDEN EAGLE II	SPEED MEASURING DEVICE	N/A	PAS
XE15018	KUSTOM SIGNALS, INC.	EAGLE II	SPEED MEASURING DEVICE	N/A	PAS:
VEIDOIO	NOOLOW SIGNALS, INC.			1	

Number Processed: 11

Date:

Customer Signature:





T.425.895.8617, F.425.702.9358

ON-SITE CERTIFICATE #: OS3-00325473

CERTIFICATE OF CALIBRATION

STANDARD CALIBRATION

CLE ELUM POLICE DEPARTMENT

807 W. 2ND STREET CLE ELUM, WA 98922

This certifies that the instrument listed herein was calibrated by Cascade Engineering Services' Calibration Laboratory, which is fully accredited in accordance with the recognized International Standards ISO/IEC 17025:2017 General Requirements for the Competence of Testing and Calibration Laboratories. Cascade Engineering Services' Calibration Laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and any additional program requirements in the field of calibration. Standards used to perform this calibration are certified by or traceable to NIST, natural physical constants, consensus standards or derived by the ratio type of calibrations. All calibrations are performed to manufacturer's specifications unless otherwise noted. Standard Calibration, while still traceable, does not meet all requirements for an Accredited Calibration per ISO/IEC 17025:2017, that is As Found data for equipment in tolerance and Measurement Uncertainties are not recorded. This certificate shall not be reproduced, except in full, without prior written approval of the laboratory.

DESCRI	PTION:	SPE	ED MEASURING DEVICE	CE		
ASSET #:		DC106052	SERIAL NUMBER:_		DC106052	
MANUFA	CTURER: AP	PLIED CONCEPTS INC.	MODEL NUMBER:_	5	STALKER DUA	L
DEPART	MENT:	N/A	LOCATION:_		N/A	
ENVIRON	MENT:	71.7 °F/36.8 %RH	BASIC ACCURACY:_	REFEREN	ICE CAL PROC	CEDURES
CAL INTE	RVAL:	12 MONTHS	DUE DATE:_		March 19, 2022)
Initial testi	ng found this equipr	ON AS RECEIVED ment to be "IN TOLERANCE", as ON AS DELIVERED				
		ration, measured values were "IN PPLIED WITH THIS DEVIC		the basic a	accuracy stated	l above.
	ESCRIPTION	SERIAL NUMBER	RATED SPEED		FREQUEN	ICY
TUNING FO		FA286716 FB394845	25	M.P.H.		2618 H
The state of the s		requency: 34.737 GHz	40	M.P.H.		4176 H
		OR CERTIFICATION				
I.D.	MODEL	MANUFACTURER	DESCR	RIPTION		DUE DATE
MET1231	VOCAR HR	DB INNOVATIONS	HAND HELD RADAR CERTIFIC	ATION SYSTE	M	08/31/2021
MET1232	VOCAR HR WAND	DB INNOVATIONS	VOCAR HR WAND			08/31/2021
PROCE	DURE(S) USED	FOR CERTIFICATION				
Barrier and the second	JMENT ID		ESCRIPTION		REV	REV DATE
CP-SMD-00	1 RADA	R AND LIDAR CALIBRATION PROCED	URE		NEW	11/14/2019
CERTIF	ICATION NOTE	S				
I certify (or	declare) under penalt	ry of perjury under the laws of the St	ate of Washington that the above	e information	is true and corre	ect
			LOCA	TION:	CLE ELUM	, WA
PERFOR	WED BY	METROLOGIST JOHN GRAY	CALIBRATION	DATE:	Friday, March	19, 2021





6640 185th Ave NE, Redmond, WA 98052 T.425.895.8617, F.425.702.9358

DESCRIPTION:

ON-SITE CERTIFICATE #: OS3-00325471

CERTIFICATE OF CALIBRATION

STANDARD CALIBRATION CLE ELUM POLICE DEPARTMENT

807 W. 2ND STREET CLE ELUM, WA 98922

This certifies that the instrument listed herein was calibrated by Cascade Engineering Services' Calibration Laboratory, which is fully accredited in accordance with the recognized International Standards ISO/IEC 17025:2017 General Requirements for the Competence of Testing and Calibration Laboratories. Cascade Engineering Services' Calibration Laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and any additional program requirements in the field of calibration. Standards used to perform this calibration are certified by or traceable to NIST, natural physical constants, consensus standards or derived by the ratio type of calibrations. All calibrations are performed to manufacturer's specifications unless otherwise noted. Standard Calibration, while still traceable, does not meet all requirements for an Accredited Calibration per ISO/IEC 17025:2017, that is "As Found" data for equipment in tolerance and Measurement Uncertainties are not recorded. This certificate shall not be reproduced, except in full, without prior written approval of the laboratory.

SPEED MEASURING DEVICE

ASSET #:		XE15010	SERIAL NUMBER:		XE15010	
MANUFAC	TURER: KL	JSTOM SIGNALS, INC.	MODEL NUMBER:		EAGLE II	
DEPARTM	ENT:	N/A	LOCATION:		202	
ENVIRON	MENT:	71.7 °F/36.8 %RH	BASIC ACCURACY:	REFEREN	NCE CAL PRO	CEDURES
CAL INTE	RVAL:	12 MONTHS	DUE DATE:		March 19, 2022	2
Initial testin EQUIPM At the com	ng found this equipment of the control of the calibration of the calib	N AS RECEIVED nent to be "IN TOLERANCE", as N AS DELIVERED ation, measured values were "IN	N TOLERANCE", as defined b			d above.
	SCRIPTION	SERIAL NUMBER	RATED SPEED		FREQUEN	ICY
TUNING FOR	RK ONE	61207	3	0 M.P.H.		3214 Hz
TUNING FOR		61442 quency: 35.473 GHz		5 M.P.H.)1, Frequency:	5904 Hz
			Antenna 2 v	JN. DLZIV	or, rrequency.	33.474 0112
		OR CERTIFICATION	DECC	DIDTION		DUE DATE
I.D. MET1231	MODEL VOCAR HR	MANUFACTURER DB INNOVATIONS	HAND HELD RADAR CERTIFIC	RIPTION	=NA	DUE DATE 08/31/2021
MET1232	VOCAR HR WAND	DB INNOVATIONS	VOCAR HR WAND	ATIONOTOTI	_ivi	08/31/2021
PROCEE	URE(S) USED	FOR CERTIFICATION				
	MENT ID		ESCRIPTION		REV	REV DATE
CP-SMD-001	RADAR	AND LIDAR CALIBRATION PROCED	URE		NEW	11/14/2019
CERTIFI	CATION NOTES					
I certify (or o	declare) under penalty	of perjury under the laws of the St	ate of Washington that the above	e information	is true and corre	ct
				ATION:	CLE ELUM	
PERFORM	IED BY	METROLOGIST-JOHN GRAY	CALIBRATION	DATE:	Friday, March	19, 2021





T.425.895.8617, F.425.702.9358

ON-SITE CERTIFICATE #: OS3-00325478

CERTIFICATE OF CALIBRATION

STANDARD CALIBRATION

CLE ELUM POLICE DEPARTMENT

807 W. 2ND STREET CLE ELUM, WA 98922

This certifies that the instrument listed herein was calibrated by Cascade Engineering Services' Calibration Laboratory, which is fully accredited in accordance with the recognized International Standards ISO/IEC 17025:2017 General Requirements for the Competence of Testing and Calibration Laboratories. Cascade Engineering Services' Calibration Laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and any additional program requirements in the field of calibration. Standards used to perform this calibration are certified by or traceable to NIST, natural physical constants, consensus standards or derived by the ratio type of calibrations. All calibrations are performed to manufacturer's specifications unless otherwise noted. Standard Calibration, while still traceable, does not meet all requirements for an Accredited Calibration per ISO/IEC 17025:2017, that is As Found data for equipment in tolerance and Measurement Uncertainties are not recorded. This certificate shall not be reproduced, except in full, without prior written approval of the laboratory.

DESCRI	IPTION:	SPE	ED MEASURING DEVIC	E		
ASSET #:		XE15012	SERIAL NUMBER:		XE15012	
MANUFA	CTURER: K	USTOM SIGNALS, INC.	MODEL NUMBER:	G	OLDEN EAGLE	II
DEPART	WENT:	N/A	LOCATION:		204	
ENVIRON	IMENT:	71.7 °F/36.8 %RH	BASIC ACCURACY:	REFERE	NCE CAL PRO	CEDURES
CAL INTE	RVAL:	12 MONTHS	DUE DATE:		March 19, 2022	2
EQUIPN At the con	ing found this equipr MENT CONDITION I pletion of the calibr	ON AS RECEIVED ment to be "IN TOLERANCE", as ON AS DELIVERED ration, measured values were "IN	TOLERANCE", as defined by			l above.
		PPLIED WITH THIS DEVIC				
TUNING FO	DESCRIPTION	SERIAL NUMBER 39203	RATED SPEED	M.P.H.	FREQUEN	3214 H
TUNING FO		41093		M.P.H.		5898 H
Antenna	1 SN: DE16732, Fre	equency: 35.483 GHz				
STANDA	ARD(S) USED F	OR CERTIFICATION				
I.D.	MODEL	MANUFACTURER	DESCR	IPTION		DUE DATE
MET1231	VOCAR HR	DB INNOVATIONS	HAND HELD RADAR CERTIFICA	ATION SYST	EM	08/31/2021
MET1232	VOCAR HR WAND	DB INNOVATIONS	VOCAR HR WAND			08/31/2021
PROCE	DURE(S) USED	FOR CERTIFICATION				
DOC	UMENT ID	DE	SCRIPTION		REV	REV DATE
CP-SMD-00	1 RADA	R AND LIDAR CALIBRATION PROCEDU	JRE		NEW	11/14/2019
	declare) under penalt	S by of perjury under the laws of the Sta	ate of Washington that the above	e information	n is true and corre	ect
		096	LOCA	TION:	CLE ELUM	. WA
PERFOR	MED BY	METROLOGIST JOHN GRAY	CALIBRATION		Friday, March	-





T.425.895.8617, F.425.702.9358

DESCRIPTION.

ON-SITE CERTIFICATE #: OS3-00325481

CERTIFICATE OF CALIBRATION

STANDARD CALIBRATION

CLE ELUM POLICE DEPARTMENT

807 W. 2ND STREET CLE ELUM, WA 98922

This certifies that the instrument listed herein was calibrated by Cascade Engineering Services' Calibration Laboratory, which is fully accredited in accordance with the recognized International Standards ISO/IEC 17025:2017 General Requirements for the Competence of Testing and Calibration Laboratories. Cascade Engineering Services' Calibration Laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and any additional program requirements in the field of calibration. Standards used to perform this calibration are certified by or traceable to NIST, natural physical constants, consensus standards or derived by the ratio type of calibrations. All calibrations are performed to manufacturer's specifications unless otherwise noted. Standard Calibration, while still traceable, does not meet all requirements for an Accredited Calibration per ISO/IEC 17025:2017, that is As Found data for equipment in tolerance and Measurement Uncertainties are not recorded. This certificate shall not be reproduced, except in full, without prior written approval of the laboratory.

CDEED MEACHDING DEVICE

DESCRI	FIION.	SPL	ED MENSORMA DEVIC	<u> </u>		
ASSET #:		XE14969	SERIAL NUMBER:		XE14969	
MANUFAC	CTURER: K	USTOM SIGNALS, INC.	MODEL NUMBER:	(GOLDEN EAGLE	II
DEPARTM	MENT:	N/A	LOCATION:		205	
ENVIRON	MENT:	71.7 °F/36.8 %RH	BASIC ACCURACY:_	REFERE	ENCE CAL PRO	CEDURES
CAL INTE	RVAL:	12 MONTHS	DUE DATE:_		March 19, 2022	2
Initial testin	ng found this equipr	ON AS RECEIVED nent to be "IN TOLERANCE", as ON AS DELIVERED ation, measured values were "IN				d above
				y the basic	c accuracy states	above.
	G FORK(S) SUI	PPLIED WITH THIS DEVIC SERIAL NUMBER	RATED SPEED		FREQUEN	ICY
TUNING FOR	RK ONE	39235	30	M.P.H.		3212 H: 5896 H:
		41070 equency: 35.484 GHz		M.P.H.	064, Frequency:	
		OR CERTIFICATION	Antenna 2 C	N. DLZI	oo4, i requericy.	33.431 011
I.D.	MODEL	MANUFACTURER	DESCE	RIPTION		DUE DATE
MET1231	VOCAR HR	DB INNOVATIONS	HAND HELD RADAR CERTIFIC		TEM	08/31/2021
MET1232	VOCAR HR WAND	DB INNOVATIONS	VOCAR HR WAND			08/31/2021
PROCE	DURE(S) USED	FOR CERTIFICATION				
	JMENT ID		SCRIPTION		REV	REV DATE
CP-SMD-001	1 RADA	R AND LIDAR CALIBRATION PROCEDI	URE		NEW	11/14/2019
	declars) under populi	S y of perjury under the laws of the St	ate of Washington that the above	o informati	on is true and corr	act
r certify (or	deciare) under penan	y or perjury under the laws of the St	ate or vvasimigion that the abov	e iiiioiiiiali	on is true and com	501
		075	LOCA	TION:	CLE ELUM	, WA
PERFORM	WED BY	METROLOGIST JOHN GRAY	CALIBRATION	DATE:	Friday, March	19, 2021
				-		





6640 185th Ave NE, Redmond, WA 98052 T.425.895.8617, F.425.702.9358

ON-SITE CERTIFICATE #: OS3-00325476

CERTIFICATE OF CALIBRATION

STANDARD CALIBRATION

CLE ELUM POLICE DEPARTMENT

807 W. 2ND STREET CLE ELUM, WA 98922

This certifies that the instrument listed herein was calibrated by Cascade Engineering Services' Calibration Laboratory, which is fully accredited in accordance with the recognized International Standards ISO/IEC 17025:2017 General Requirements for the Competence of Testing and Calibration Laboratories. Cascade Engineering Services' Calibration Laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and any additional program requirements in the field of calibration. Standards used to perform this calibration are certified by or traceable to NIST, natural physical constants, consensus standards or derived by the ratio type of calibrations. All calibrations are performed to manufacturer's specifications unless otherwise noted. Standard Calibration, while still traceable, does not meet all requirements for an Accredited Calibration per ISO/IEC 17025:2017, that is As Found data for equipment in tolerance and Measurement Uncertainties are not recorded. This certificate shall not be reproduced, except in full, without prior written approval of the laboratory.

DESCRI	PTION:	SPE	ED MEASURING DEVI	CE	Carried Advisor on	
ASSET #:		PYT846001765	SERIAL NUMBER:		PYT846001765	;
MANUFAC	CTURER:	MPH INDUSTRIES	MODEL NUMBER:		PYTHON III	
DEPART	MENT:	N/A	LOCATION:		206	
ENVIRON	MENT:	71.7 °F/36.8 %RH	BASIC ACCURACY:	REFERE	NCE CAL PRO	CEDURES
CAL INTE	RVAL:	12 MONTHS	DUE DATE:		March 19, 2022	<u>}</u>
		N AS RECEIVED nent to be "IN TOLERANCE", as	defined by the basic accura	cv stated a	bove.	
EQUIPM	IENT CONDITIO	N AS DELIVERED ation, measured values were "IN				d above.
		PPLIED WITH THIS DEVIC				
TUNING FO	ESCRIPTION	SERIAL NUMBER 854221	RATED SPEED	85 M.P.H.	FREQUEN	1CY 2522 H
TUNING FO		854244		55 M.P.H.		4680 H
Antenna 1	1 SN: PYT83100213	7, Frequency: 24.191 GHz				
STANDA	ARD(S) USED FO	OR CERTIFICATION				
I.D.	MODEL	MANUFACTURER	DESCI	RIPTION		DUE DATE
MET1231	VOCAR HR	DB INNOVATIONS	HAND HELD RADAR CERTIFIC	CATION SYST	ГЕМ	08/31/2021
MET1232	VOCAR HR WAND	DB INNOVATIONS	VOCAR HR WAND			08/31/2021
PROCE	DURE(S) USED	FOR CERTIFICATION				
DOCL	JMENT ID	DE	SCRIPTION		REV	REV DATE
CP-SMD-00	1 RADAF	R AND LIDAR CALIBRATION PROCEDU	URE		NEW	11/14/2019
CERTIF	ICATION NOTES	5				
I certify (or	declare) under penalty	y of perjury under the laws of the Sta	ate of Washington that the abov	e information	on is true and corre	ect
			LOCA	ATION:	CLE ELUM	, WA
PERFOR	MED BY	METROLOGIST-JOHN GRAY	CALIBRATION	DATE:	Friday, March	19, 2021





T.425.895.8617, F.425.702.9358

ON-SITE CERTIFICATE #: OS3-00325480

CERTIFICATE OF CALIBRATION

STANDARD CALIBRATION

CLE ELUM POLICE DEPARTMENT

807 W. 2ND STREET CLE ELUM, WA 98922

This certifies that the instrument listed herein was calibrated by Cascade Engineering Services' Calibration Laboratory, which is fully accredited in accordance with the recognized International Standards ISO/IEC 17025:2017 General Requirements for the Competence of Testing and Calibration Laboratories. Cascade Engineering Services' Calibration Laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and any additional program requirements in the field of calibration. Standards used to perform this calibration are certified by or traceable to NIST, natural physical constants, consensus standards or derived by the ratio type of calibrations. All calibrations are performed to manufacturer's specifications unless otherwise noted. Standard Calibration, while still traceable, does not meet all requirements for an Accredited Calibration per ISO/IEC 17025:2017, that is As Found data for equipment in tolerance and Measurement Uncertainties are not recorded. This certificate shall not be reproduced, except in full, without prior written approval of the laboratory.

DESCRI	PTION:	SPE	ED MEASURING DEVICE		
ASSET #:		XE08339	SERIAL NUMBER:	XE08339	
MANUFAC	TURER: KU	STOM SIGNALS, INC.	MODEL NUMBER:	GOLDEN EAGLE	E II
DEPARTM	IENT:	N/A	LOCATION:	207	
ENVIRON	MENT:	71.7 °F/36.8 %RH	BASIC ACCURACY: REF	ERENCE CAL PRO	CEDURES
CAL INTER	RVAL:	12 MONTHS	DUE DATE:	March 19, 2022	2
Initial testin			defined by the basic accuracy state	ed above.	
At the com	pletion of the calibrat	ion, measured values were "IN	TOLERANCE", as defined by the b	pasic accuracy stated	d above.
TUNNING	G FORK(S) SUPI	PLIED WITH THIS DEVIC	E		
	ESCRIPTION	SERIAL NUMBER	RATED SPEED	FREQUEN	
TUNING FOR		39205 41098	30 M.P.H 55 M.P.H		3216 Hz 5904 Hz
Antenna 1	SN: DE27085, Freq	uency: 35.486 GHz			
STANDA	(RD(S) USED FO	R CERTIFICATION			
STANDA I.D.	ARD(S) USED FO MODEL	R CERTIFICATION MANUFACTURER	DESCRIPTIO	ON	DUE DATE
I.D. MET1231	MODEL VOCAR HR	MANUFACTURER DB INNOVATIONS	HAND HELD RADAR CERTIFICATION		08/31/2021
I.D.	MODEL	MANUFACTURER			
I.D. MET1231 MET1232	MODEL VOCAR HR VOCAR HR WAND	MANUFACTURER DB INNOVATIONS	HAND HELD RADAR CERTIFICATION		08/31/2021
I.D. MET1231 MET1232 PROCED DOCU	MODEL VOCAR HR VOCAR HR WAND DURE(S) USED F	MANUFACTURER DB INNOVATIONS DB INNOVATIONS OR CERTIFICATION	HAND HELD RADAR CERTIFICATION		08/31/2021 08/31/2021
I.D. MET1231 MET1232 PROCED	MODEL VOCAR HR VOCAR HR WAND OURE(S) USED F	MANUFACTURER DB INNOVATIONS DB INNOVATIONS OR CERTIFICATION	HAND HELD RADAR CERTIFICATION OF VOCAR HR WAND	SYSTEM	08/31/2021 08/31/2021
I.D. MET1231 MET1232 PROCED DOCU CP-SMD-001 CERTIFIC	MODEL VOCAR HR VOCAR HR WAND DURE(S) USED F MENT ID RADAR CATION NOTES declare) under penalty	MANUFACTURER DB INNOVATIONS DB INNOVATIONS FOR CERTIFICATION DE AND LIDAR CALIBRATION PROCEDI	HAND HELD RADAR CERTIFICATION OF VOCAR HR WAND	REV NEW mation is true and corre	08/31/2021 08/31/2021 REV DATE 11/14/2019 ect





T.425.895.8617, F.425.702.9358

ON-SITE CERTIFICATE #: OS3-00325479

CERTIFICATE OF CALIBRATION

STANDARD CALIBRATION

CLE ELUM POLICE DEPARTMENT

807 W. 2ND STREET CLE ELUM, WA 98922

This certifies that the instrument listed herein was calibrated by Cascade Engineering Services' Calibration Laboratory, which is fully accredited in accordance with the recognized International Standards ISO/IEC 17025:2017 General Requirements for the Competence of Testing and Calibration Laboratories. Cascade Engineering Services' Calibration Laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and any additional program requirements in the field of calibration. Standards used to perform this calibration are certified by or traceable to NIST, natural physical constants, consensus standards or derived by the ratio type of calibrations. All calibrations are performed to manufacturer's specifications unless otherwise noted. Standard Calibration, while still traceable, does not meet all requirements for an Accredited Calibration per ISO/IEC 17025:2017, that is As Found data for equipment in tolerance and Measurement Uncertainties are not recorded. This certificate shall not be reproduced, except in full, without prior written approval of the laboratory.

DESCRI	PTION:	SPE	ED MEASURING DEVICE		
ASSET #:		KK38300	SERIAL NUMBER:	KK38300	
MANUFAC	TURER: KI	USTOM SIGNALS, INC.	MODEL NUMBER:	TROOPER	
DEPARTM	IENT:	N/A	LOCATION:	209	
ENVIRONI	MENT:	71.7 °F/36.8 %RH	BASIC ACCURACY: REFE	RENCE CAL PRO	CEDURES
CAL INTE	RVAL:	12 MONTHS	DUE DATE:	March 19, 2022	2
EQUIPM At the com	ng found this equipment of the condition of the calibration of the cal	ON AS DELIVERED ation, measured values were "IN PPLIED WITH THIS DEVIC	defined by the basic accuracy state TOLERANCE", as defined by the b	asic accuracy stated	
	ESCRIPTION	SERIAL NUMBER	RATED SPEED	FREQUEN	AND DESCRIPTION OF THE PARTY OF
TUNING FOR		26717 25548	35 M.P.H. 65 M.P.H.		2544 Hz 4730 Hz
Antenna 1	SN: CC39303, Fre	equency: 24.168 GHz			
STANDA	RD(S) USED F	OR CERTIFICATION			
I.D.	MODEL	MANUFACTURER	DESCRIPTIO	N	DUE DATE
MET1231	VOCAR HR	DB INNOVATIONS	HAND HELD RADAR CERTIFICATION S	SYSTEM	08/31/2021
MET1232	VOCAR HR WAND	DB INNOVATIONS	VOCAR HR WAND		08/31/2021
PROCE	DURE(S) USED	FOR CERTIFICATION			
DOCU		DF	SCRIPTION	REV	REV DATE
CP-SMD-001	MENTID	DL	JOINT HON	11=	
CP-51VID-001		R AND LIDAR CALIBRATION PROCEDU			11/14/2019
CERTIFI	RADAI	R AND LIDAR CALIBRATION PROCEDU		NEW	11/14/2019
CERTIFI	RADAI	R AND LIDAR CALIBRATION PROCEDU	JRE	NEW	11/14/2019 ect





T.425.895.8617, F.425.702.9358

ON-SITE CERTIFICATE #: OS3-00325474

CERTIFICATE OF CALIBRATION

STANDARD CALIBRATION

CLE ELUM POLICE DEPARTMENT

807 W. 2ND STREET CLE ELUM, WA 98922

This certifies that the instrument listed herein was calibrated by Cascade Engineering Services' Calibration Laboratory, which is fully accredited in accordance with the recognized International Standards ISO/IEC 17025:2017 General Requirements for the Competence of Testing and Calibration Laboratories. Cascade Engineering Services' Calibration Laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and any additional program requirements in the field of calibration. Standards used to perform this calibration are certified by or traceable to NIST, natural physical constants, consensus standards or derived by the ratio type of calibrations. All calibrations are performed to manufacturer's specifications unless otherwise noted. Standard Calibration, while still traceable, does not meet all requirements for an Accredited Calibration per ISO/IEC 17025:2017, that is As Found data for equipment in tolerance and Measurement Uncertainties are not recorded. This certificate shall not be reproduced, except in full, without prior written approval of the laboratory.

DESCR	IPTION:	SPE	ED MEASURING DEVIC	E		
ASSET #:		XE08340	SERIAL NUMBER:		XE08340	
MANUFA	CTURER: KI	JSTOM SIGNALS, INC.	MODEL NUMBER:	GOL	DEN EAGLE	: 11
DEPART	MENT:	N/A	LOCATION:		210	
ENVIRON	IMENT:	71.7 °F/36.8 %RH	BASIC ACCURACY:	REFERENCI	E CAL PRO	CEDURES
CAL INTE	RVAL:	12 MONTHS	DUE DATE:	Ma	rch 19, 2022	2
EQUIPN At the con	ing found this equipn INFNT CONDITION TO THE CALIBRATE TH	N AS RECEIVED nent to be "IN TOLERANCE", as N AS DELIVERED ation, measured values were "IN PPLIED WITH THIS DEVIC	TOLERANCE", as defined by			l above.
	DESCRIPTION	SERIAL NUMBER	RATED SPEED		FREQUEN	ICY
TUNING FO		32821 34576		M.P.H. M.P.H.		3218 H 5900 H
		quency: 35.474 GHz	33	IVI.F.J I.		33001
		OR CERTIFICATION				
I.D.	MODEL	MANUFACTURER	DESCR	IPTION		DUE DATE
MET1231	VOCAR HR	DB INNOVATIONS	HAND HELD RADAR CERTIFICA	TION SYSTEM		08/31/2021
MET1232	VOCAR HR WAND	DB INNOVATIONS	VOCAR HR WAND			08/31/2021
PROCE	DURE(S) USED	FOR CERTIFICATION				
DOC	UMENT ID	DE	SCRIPTION			REV DATE
CP-SMD-00	1 RADAF	R AND LIDAR CALIBRATION PROCEDI	URE		NEW	11/14/2019
	declare) under penalt	y of perjury under the laws of the Sta	ate of Washington that the above	e information is	true and corn	ect
, (-						
			LOCA	TION:	CLE ELUM	, WA
		(XA				
PERFOR	MED BY	METROLOGIST JOHN GRAY	CALIBRATION	DATE: Fr	iday, March	19, 2021



THE CHART

6640 185th Ave NE, Redmond, WA 98052

T.425.895.8617, F.425.702.9358

ON-SITE CERTIFICATE #: OS3-00325472

CERTIFICATE OF CALIBRATION

STANDARD CALIBRATION

CLE ELUM POLICE DEPARTMENT

807 W. 2ND STREET CLE ELUM, WA 98922

This certifies that the instrument listed herein was calibrated by Cascade Engineering Services' Calibration Laboratory, which is fully accredited in accordance with the recognized International Standards ISO/IEC 17025:2017 General Requirements for the Competence of Testing and Calibration Laboratories. Cascade Engineering Services' Calibration Laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and any additional program requirements in the field of calibration. Standards used to perform this calibration are certified by or traceable to NIST, natural physical constants, consensus standards or derived by the ratio type of calibrations. All calibrations are performed to manufacturer's specifications unless otherwise noted. Standard Calibration, while still traceable, does not meet all requirements for an Accredited Calibration per ISO/IEC 17025:2017, that is As Found data for equipment in tolerance and Measurement Uncertainties are not recorded. This certificate shall not be reproduced, except in full, without prior written approval of the laboratory.

DESCRI	PTION:	SPE	ED MEASURING DEVICE			
ASSET #:		XE15018	SERIAL NUMBER:	XE15018		
MANUFAC	CTURER: KU	STOM SIGNALS, INC.	MODEL NUMBER:	EAGLE II		
DEPART	MENT:	N/A	LOCATION:	225		
ENVIRON	MENT:	71.7 °F/36.8 %RH	BASIC ACCURACY: REFER	RENCE CAL PRO	E CAL PROCEDURES	
CAL INTE	RVAL:	12 MONTHS	DUE DATE:	March 19, 202	2	
EQUIPN At the com	IENT CONDITION	ent to be "IN TOLERANCE", as N AS DELIVERED tion, measured values were "IN	defined by the basic accuracy stated TOLERANCE", as defined by the ba		d above.	
	ESCRIPTION	PLIED WITH THIS DEVIC SERIAL NUMBER	RATED SPEED	FREQUE	NCY	
TUNING FO	RK ONE	32829	30 M.P.H.		3216 H	
TUNING FO	RK TWO 1 SN: DE16734, Freq	34550	55 M.P.H. 5904 Antenna 2 SN: DE27084, Frequency: 35.484 GI			
			Antenna 2 ON. DE2	. roo4, r requeries	. 00.404 01	
I.D.	MODEL	R CERTIFICATION MANUFACTURER	DESCRIPTION		DUE DAT	
MET1231	VOCAR HR	DB INNOVATIONS	HAND HELD RADAR CERTIFICATION SY		08/31/2021	
MET1232	VOCAR HR WAND	DB INNOVATIONS	VOCAR HR WAND		08/31/2021	
PROCE	DURE(S) USED F	OR CERTIFICATION				
	JMENT ID		SCRIPTION	REV	REV DAT	
CP-SMD-00		AND LIDAR CALIBRATION PROCEDU	URE	NEW	11/14/2019	
	declare) under penalty		ate of Washington that the above informa	ation is true and cor	rect	
		075	LOCATION:	CLE ELUM	1, WA	
PERFORI	MED DV	X	CALIBRATION DATE:	Friday, March	10 2021	





T.425.895.8617, F.425.702.9358

DESCRIPTION:

ON-SITE CERTIFICATE #: OS3-00325475

CERTIFICATE OF CALIBRATION

STANDARD CALIBRATION

CLE ELUM POLICE DEPARTMENT

807 W. 2ND STREET CLE ELUM, WA 98922

This certifies that the instrument listed herein was calibrated by Cascade Engineering Services' Calibration Laboratory, which is fully accredited in accordance with the recognized International Standards ISO/IEC 17025:2017 General Requirements for the Competence of Testing and Calibration Laboratories. Cascade Engineering Services' Calibration Laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and any additional program requirements in the field of calibration. Standards used to perform this calibration are certified by or traceable to NIST, natural physical constants, consensus standards or derived by the ratio type of calibrations. All calibrations are performed to manufacturer's specifications unless otherwise noted. Standard Calibration, while still traceable, does not meet all requirements for an Accredited Calibration per ISO/IEC 17025:2017, that is As Found data for equipment in tolerance and Measurement Uncertainties are not recorded. This certificate shall not be reproduced, except in full, without prior written approval of the laboratory.

SPEED MEASURING DEVICE

ASSET #:		FF12409	SERIAL NUMBER:		FF12409	
MANUFAC	TURER: KU	STOM SIGNALS, INC.	MODEL NUMBER:	FALCON		
DEPARTM	ENT:	N/A	LOCATION:	PA	ATROL - HANDH	ELD
ENVIRON	MENT:	71.7 °F/36.8 %RH	BASIC ACCURACY:	REFERE	ENCE CAL PRO	CEDURES
CAL INTE	RVAL:	12 MONTHS	DUE DATE:		March 19, 2022	
						-22-53-0-0-0-0-0
EQUIPM	ENT CONDITION	AS RECEIVED				
Initial testin	ng found this equipme	ent to be "IN TOLERANCE", a	s defined by the basic accurac	y stated a	bove.	
		AS DELIVERED				
At the com	pletion of the calibrat	tion, measured values were "I	N TOLERANCE", as defined by	y the basic	c accuracy stated	above.
TUNNING	G FORK(S) SUPI	PLIED WITH THIS DEVI	CE			
DI	SCRIPTION	SERIAL NUMBER	RATED SPEED		FREQUEN	ICY
TUNING FOR		26500	50	M.P.H.		3648 Hz
TUNING FOR		N/A				
		requency: 24.156 GHz				
STANDA	RD(S) USED FO	R CERTIFICATION				
I.D.	MODEL	MANUFACTURER	DESCR	RIPTION		DUE DATE
MET1231	VOCAR HR	DB INNOVATIONS	HAND HELD RADAR CERTIFIC	ATION SYS	TEM	08/31/2021
MET1232	VOCAR HR WAND	DB INNOVATIONS	VOCAR HR WAND			08/31/2021
PROCE	OURE(S) USED F	OR CERTIFICATION				
DOCU	MENT ID	D	ESCRIPTION		REV	REV DATE
CP-SMD-001	RADAR	AND LIDAR CALIBRATION PROCED	DURE		NEW	11/14/2019
CERTIFI	CATION NOTES					
I certify (or o	declare) under penalty	of periury under the laws of the S	State of Washington that the abov	e informati	on is true and corr	ect
, (,,		3			
			LOCA	TION:	CLE ELUM	, WA
PERFORM		VA				
. Litt Oitti	IED BY	METROLOGIST: JOHN GRAY	CALIBRATION	DATE:	Friday, March	19, 2021





T.425.895.8617, F.425.702.9358

ON-SITE CERTIFICATE #: OS3-00325477

CERTIFICATE OF CALIBRATION

STANDARD CALIBRATION

CLE ELUM POLICE DEPARTMENT

807 W. 2ND STREET CLE ELUM, WA 98922

This certifies that the instrument listed herein was calibrated by Cascade Engineering Services' Calibration Laboratory, which is fully accredited in accordance with the recognized International Standards ISO/IEC 17025:2017 General Requirements for the Competence of Testing and Calibration Laboratories. Cascade Engineering Services' Calibration Laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and any additional program requirements in the field of calibration. Standards used to perform this calibration are certified by or traceable to NIST, natural physical constants, consensus standards or derived by the ratio type of calibrations. All calibrations are performed to manufacturer's specifications unless otherwise noted. Standard Calibration, while still traceable, does not meet all requirements for an Accredited Calibration per ISO/IEC 17025:2017, that is As Found data for equipment in tolerance and Measurement Uncertainties are not recorded. This certificate shall not be reproduced, except in full, without prior written approval of the laboratory.

DESCRIP	PTION:	SPE	ED MEASURING DEVI	CE			
ASSET #: AA11249 MANUFACTURER: KUSTOM SIGNALS, INC.		SERIAL NUMBER:					
		MODEL NUMBER:		HR-12			
DEPARTM	ENT:	N/A	LOCATION:	PAT	ROL - HANDH	ELD	
ENVIRONN	MENT:	71.7 °F/36.8 %RH	BASIC ACCURACY:	REFEREN	RENCE CAL PROCEDURES		
CAL INTER	RVAL:	12 MONTHS	DUE DATE:		March 19, 2022		
EQUIPMI At the comp	g found this equipme ENT CONDITION pletion of the calibrat	I AS RECEIVED ent to be "IN TOLERANCE", as I AS DELIVERED tion, measured values were "IN PLIED WITH THIS DEVIC	TOLERANCE", as defined by			d above.	
	SCRIPTION	SERIAL NUMBER	RATED SPEED		FREQUEN	THE RESERVE OF THE PARTY OF THE	
TUNING FOR TUNING FOR		292168 24719		5 M.P.H. 5 M.P.H.		2538 Hz 4730 Hz	
0120213340000000000000000000000000000000		requency: 24.184 GHz					
		R CERTIFICATION					
I.D.	MODEL	MANUFACTURER	DESCR	RIPTION		DUE DATE	
MET1231	VOCAR HR	DB INNOVATIONS	HAND HELD RADAR CERTIFIC	ATION SYSTE	EM	08/31/2021	
MET1232	VOCAR HR WAND	DB INNOVATIONS	VOCAR HR WAND			08/31/2021	
PROCED	URE(S) USED F	OR CERTIFICATION					
DOCU	MENT ID	DE	ESCRIPTION		REV	REV DATE	
CP-SMD-001	RADAR	AND LIDAR CALIBRATION PROCEDI	URE		NEW	11/14/2019	
	CATION NOTES	of perjury under the laws of the St	ate of Washington that the abov	e information	n is true and corr	ect	
		096	1004	ATION:	CLE ELUM	WA	
PERFORM	ED BY	XA	CALIBRATION		Friday, March		

THIS LABORATORY IS A2LA ACCREDITED TO ISO/IEC 17025:2017 (GENERAL REQUIREMENTS FOR THE COMPETENCE OF TESTING AND CALIBRATION LABORATORIES). CERTIFICATE #: 2560.01 F540-1.3/REVISION M Page 1 of 1 WASHINGTON STATE PATROL CRIMINAL RECORDS DIVISION P.O. BOX 42633 (360) 534-2000 WWW.WSP.WA.GO

CRIMINAL HISTORY INFORMATION AS OF 03/09/2020

THE FOLLOWING TRANSCRIPT OF RECORD IS FURNISHED FOR OFFICIAL USE ONLY. SECONDARY DISSEMINATION OF THIS CRIMINAL HISTORY RECORD INFORMATION IS PROHIBITED UNLESS IN COMPLIANCE WITH THE WASHINGTON STATE CRIMINAL RECORDS PRIVACY ACT, CHAPTER 10.97 RCW. CONTINUE WITH THE WASHINGTHE STATE CRIMINAL RECORD THAT WASHINGTHE OF A NON-CRIMINAL USTICE AGENCIES WHO HAVE STATUTORY AUTHORITY TO RECEIVE NON-CONVICTION CRIMINAL HISTORY RECORDS MAY NOT USE SEALED JUVENILE RECORD INFORMATION IN MAKING LICENSING OR EMPLOYMENT SUITABILITY DETERMINATIONS. POSITIVE IDENTIFICATION CAN ONLY BE BASED UPON FINGERPRINT COMPARISON. BECAUSE ADDITIONS

OR DELETIONS MAY BE MADE AT ANY TIME, A NEW COPY SHOULD BE REQUESTED FOR SUBSEQUENT USE. WHEN EXPLANATION OF A CHARGE OR DISPOSITION IS NEEDED, COMMUNICATE DIRECTLY WITH THE AGENCY THAT SUPPLIED THE INFORMATION TO THE WASHINGTON STATE PATROL

A DISPOSITION STATUS OF "WARRANT ISSUED" MAY NOT INDICATE THE PRESENCE OF AN ACTIVE WARRANT, ALL WARRANT INFORMATION SHOULD BE OBTAINED FROM THE ACCESS SYSTEM AND MUST BE CONFIRMED WITH THE ENTERING AGENCY.

MASTER INFORMATION

NAME: TRAINING, RECORD SID NUMBER: WA10000825

DO8: 12/08/1975 FBI NUMBER: 91007700

DOC NUMBER: 999999

PERSON INFORMATION

DATE REPORTED DECEASED REPORTING AGENCY: DECEASED TYPE:

04/11/2014 WA0340000 REPORTED

WEIGHT EYES PLACE OF BIRTH SEX RACE HEIGHT 150 BLU BLU

WA us

CITIZENSHIE

OTHER NAMES USED OTHER DATES OF EXAMPLE ,MARY

SOC SEC

MISC NUMBER

05/15/1995 DNA TAKEN: Y DNA TYPED: Y

DLO: WSP CRIME LABATORY - SEATTLE, CODIS UNIT (206) 262-6020, STR, 105-0356X COMMENT: TWIN, TRUE NAME UNKNOWN

PHOTOGRAPH



			-		
ECADE	BAADVE	TATTOOC	ANID	ASSOLIT	ATIONS

LOCATION DESCRIPTION DESCRIPTION

CONVICTION AND/OR ADVERSE FINDING SUMMARY 5 FELONYS
MURDER-2
4 GROSS MISDEMEANORS

DISPOSITION DATE 05/15/2017 12/31/2019

DRIVE W/LICENSE SUSP OR REVOKED-3 2 CLASSIFICATIONS UNKNOWN DRUG RELATED CHARGE

12/31/2019 12/31/2019

DOC SUMMARY

RAPE OF CHILD-3

2

COMMITMENT

CLASS A FELONY

01/14/2016

CRIMINAL HISTORY RECORD INFORMATION

DATE OF ARREST: 07/01/2013 3 NAME USED: PIED, PIPER CONTRIBUTING AGENCY: WA0340000 THURSTON COUNTY SHERIFF LOCAL ID: 456789

ARREST INFORMATION:

RESIDENTIAL BURGLARY RCW: 9A.52.025(2) CLASS 8 FELONY OIN: 456789 COURT CASE NO:

DATE OF OFFENSE: 07/01/2013

ORIGINATING AGENCY: WA0340000 THURSTON COUNTY SHERIFF DISPO RESPONSIBILITY: WA034123) THURSTON COUNTY SUPERIOR COURT

POSSESS BURGLARY TOOLS RCW: 9A.52.060(2) GROSS MISDEMEANOR OIN: 456789

DATE OF OFFENSE: 07/01/2013

OUN: 306789 COURT CASE NO: ORICINATING AGENCY: WA0340000 THURSTON COUNTY SHERIFF DISPO RESPONSIBILITY: WA0341233 THURSTON COUNTY SUPERIOR COURT

COURT CASE NO: 1310069XX CONTRIBUTOR OR DISPOSITION RESPONSIBLE AGENCY: WA034123J THURSTON COUNTY SUPERIOR COURT

RESIDENTIAL BURGLARY DISPOSITION: GUILTY RCW: 9A.52.025(2) CRITERION OFFENSE

DISPOSITION DATE: 07/01/2015

SENT. DESC.: DNA: 100, COURT COSTS: 200, PRISON:24 MOS, SUPV: 12 MOS, VICTIM ASSESS:500

SUBSEQUENT DISPOSITION; DISMISSED

DATE: 11/15/2014

ORI: WA034123J

SUBSEQUENT ACTIVITY: RIGHT TO POSSESS A FIREARM RESTORED DATE: 10/31/2015

ORT: WADS41231 THURSTON COUNTY SUPERTOR COURT COMMENT: 9.41.040(4)



RAPsheet Changes Quick Reference Guide

New RAPsheet Coming May 11, 2021

What's Changing?

The new RAPsheet will have a more logical flow of information with a format that is easier to read.

How do I learn more?

Everyone is welcome to attend the "What's New with W2" or "RAPsheet" training. Sign up HERE!

What's New:

- Bold text and colored headers for easier comprehension
- New: Arrestee photograph
- New field: TCN field to track the arrest through disposition 3.
- Historical tracking number (the PCN) is captured in the
- Linear flow of arrest information 5.
- Linear flow of disposition information 6.
- New field: Subsequent Disposition lists the status of the court case. New interim disposition values are listed on the back of this QBR. Training is also available to learn
- Renamed field: Subsequent Activity is where related court orders are located. (previously called Subsequent
- Updated address format. Non-conviction RAPsheet shows both arrest and disposition information
- 10. Other Offender Registrations: formerly Monitored Populations Tracking Information
- 11. New Fingerprinting Type: Compromised Identity

ADDRESSES

RESIDENCE:
7539 S DILLARD AVE
CONCRETE, NA 98237 COUNTY: SKAGIT
DATE SUBMITTED: 07/09/2014
BEGIN DATE: 07/09/2014

COUNTRY: US

MAILING:
PO BOX: 834
CONCRETE, NA 98237
COUNTY: SKAGIT
DATE SUBMITTED: 07/09/2014
BEGIN DATE: 07/09/2014

COUNTRY: US

SCHOOL:
BATES TECHINICAL COLLEGE
1101 S YAKIMA AVE
TACOMA, MA 384057 COUNTY: PIERCE
DATE SUBMITTED: 07/09/2014
BEGIN DATE: 07/09/2014

COUNTRY: US

1

9

OTHER OFFENDER REGISTRATIONS

TYPE: FELONY FIREARMS OFFENDER REGISTRATION

REG REQ DATE: 08/13/2018 PROJECTED EXP DATE: 10/13/2022 NAME USED: BAD, HABITS

CONTRIBUTING AGENCY: WA0280000 SAN JUAN COUNTY SHERIFF

APPLICANT INFORMATION

TYPE: CRIMINAL JUSTICE

DATE: 04/20/2000

PERSONAL IDENTIFICATION

FINGERPRINT REASON: COMPROMISED IDENTITY CLAIM LOCAL ID: 999999

APPLICANT INFORMATION

ADMINISTRATIVE OFFICE OF THE COURTS	HTTP://WWW.COURTS.WA.GOV
WSP CHRU	CRIMHIS@WSP.WA.GOV OR (360) 534-2000
RECORDS IMPROVEMENT UNIT	RECORDSIMPROVEMENT@WSP.WA.GOV



Cle Elum – Roslyn – South Cle Elum Police Department

In The Heart of the Cascades

Chief of Police – Kirk Bland

807 W Second St ♦ Cle Elum, WA 98922 ♦ 509-674-2991 ♦ Fax 509-674-2918 ♦ Dispatch 509-925-8534 policerecords@cityofcleelum.com

[] 201 - Kustom Eagle II (ser #XE15010) (30 MPH Fork #61207) (55MPH Fork #61442)
[] 202 - Stalker Dual(ser #DC106052) (25 MPH Fork #FA286716) (40 MPH Fork #FB394845)
{] 204 - Golden Eagle II (ser #XE15012) (30 MPH Fork #39203) (55 MPH Fork #41093)
[] 206 - Kustom Golden Eagle II (ser #XE14969) (30MPH Fork #39235) (55MPH Fork #41070)
[] 207 - Golden Eagle II (Ser #08339) (30 MPH Fork #39205) (55 MPH Fork #41098)
[] 208 - MPH Python III (ser #PYT846001765) (35MPH Fork #854221) (65MPH Fork #854244)
[] 209 - Kustom Trooper (ser #KK38300) (35MPH Fork #26717) (65MPH Fork #25548)
[] 210 - Golden Eagle II (ser #XE08340) (30 MPH Fork #32821) (55 MPH Fork #34576)
[] 225 - Kustom Eagle II (ser #XE15018)(30 MPH Fork #32839) (55 MPH Fork #34550)
[] Falcon (HH) - (ser #FF-12409) (50MPH Fork #26500)